

IV2-008

C-peptide Controls A & B

Intended Use

C-peptide Controls A & B are intended to be used in conjunction with the Invitron C-peptide Assay (IV2-004) to provide an indication of routine assay performance.

Reagents

C-peptide Controls A & B are lyophilised samples containing low (A) and high (B) concentrations of synthetic human C-peptide in a serum matrix.

Refer to the certificate of analysis for target control values.

Each laboratory should establish its own expected concentration range.

Storage and Stability

The controls are stable for one year from the time of manufacture provided they are stored at 2-8°C. Once reconstituted, the controls are stable for up to 1 week when stored at 2-8°C.

Directions for Use

Reconstitute each of the controls by the addition of 1 ml of deionised water. Allow these to stand for 5 minutes, then mix gently to ensure all solids are dissolved.

Limitations

- Controls should not be used beyond their expiry date.
- Controls are not intended for use as calibrators.

This product contains no human derived material.

For additional information please contact:

Invitron Limited
Wyastone Business Park
Wyastone Leys
Monmouth NP25 3SR

Tel: +44 (0)1600 891536
Fax: +44 (0)1600 891537
Email: info@invitron.co.uk
www.invitron.co.uk